

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3193
 Lab File ID: VX047981.D Date Analyzed: 10/03/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:35
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	131937	5.53	225878	6.74	209083	10.04
UPPER LIMIT	263874	6.031	451756	7.239	418166	10.537
LOWER LIMIT	65968.5	5.031	112939	6.239	104542	9.537
EPA SAMPLE NO.						
RW10-MW01D-20250923	131753	5.54	267351	6.75	277307	10.04
EB04-20250923	129971	5.54	263211	6.75	269714	10.04
VX1003WBL01	132031	5.54	278145	6.75	290715	10.04
VX1003WBS02	126673	5.54	223242	6.75	203303	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.