

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3193</u>
Lab File ID:	<u>VX047896.D</u>	Date Analyzed:	<u>09/30/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:20</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	121307	12.00€				
UPPER LIMIT	242614	12.50€				
LOWER LIMIT	60653.5	11.50€				
EPA SAMPLE NO.						
RE137-20250922	124454	12.01				
RW4-20250922	125109	12.01				
RW10A-MW01I-20250923	134429	12.01				
RW8-MW01S-20250923	134231	12.01				
TT172S1-20250923	137318	12.01				
DUP12-20250923	144063	12.01				
RW10A-MW01S-20250923	128891	12.01				
RW8-MW01D1-20250923	142118	12.01				
DUP11-20250923	135633	12.01				
TT161S1-20250923	138945	12.01				
RE103D1-20250923	142951	12.01				
VX0930WBL01	133087	12.01				
VX0930WBS01	116962	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.