

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3193</u>
Lab File ID:	<u>VX047925.D</u>	Date Analyzed:	<u>10/01/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>08:44</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	111244	12.00€				
UPPER LIMIT	222488	12.50€				
LOWER LIMIT	55622	11.50€				
EPA SAMPLE NO.						
RW10-MW01D-20250923MS	115324	12.01				
RW10-MW01D-20250923MSD	109020	12.01				
TT149S1-20250922	132812	12.01				
RE137-20250922DL	119740	12.01				
RW4-20250922DL	121354	12.01				
GM38-RW3-MW1-20250922	149951	12.01				
TT20551-20250923	141662	12.01				
RE103D1-20250923DL	148894	12.01				
MW203D-20250923	151044	12.01				
TT174I1-20250924	123424	12.01				
TT162S1-20250924	118778	12.01				
RE120D3-20250924	154786	12.01				
VX1001WBL01	149007	12.01				
VX1001WBS01	117338	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.