

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3194

Client ID : SSTDCCC0.4

Date Analyzed: 10/06/2025

Lab File ID: BN038004.D

Time Analyzed: 11:06

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8577	7.804	24563	10.61	12995	14.46
UPPER LIMIT	17154	8.304	49126	11.105	25990	14.962
LOWER LIMIT	4288.5	7.304	12281.5	10.105	6497.5	13.962
EPA SAMPLE NO.						
01 PB169903BL	8521	7.82	21401	10.62	9560	14.46
02 RW11-MW01S-20250922	8149	7.80	22929	10.61	11897	14.46
03 TT189D1-20250922	8295	7.80	24176	10.61	12804	14.46
04 RW11-MW01S-20250922MS	7454	7.80	19259	10.61	8854	14.46
05 RW11-MW01S-20250922MSD	7368	7.80	19533	10.61	9117	14.46
06 TT150S1-20250922	8753	7.80	24921	10.61	12370	14.46
07 RW8-MW01D2-20250922	8364	7.80	23695	10.61	12578	14.46
08 RW8-MW01D3-20250922	8647	7.80	23227	10.61	11318	14.46
09 PB169903BS	8454	7.80	22777	10.61	10545	14.46
10 TT163S1-20250924	8745	7.80	23793	10.60	12365	14.46
11 TT188S1-20250922	8436	7.80	23277	10.61	11715	14.46
12 TT192D2-20250922	7711	7.80	21865	10.61	11194	14.46
13 TT192D1-20250922	7823	7.80	22788	10.61	10971	14.46
14 DUP10-20250922	8052	7.80	23897	10.61	12790	14.46
15 RW11-MW01I-20250922	8350	7.80	23997	10.61	12038	14.46

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.