

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3194

Client ID: SSTDCCC0.4

Date Analyzed: 10/06/2025

Lab File ID: BN038004.D

Time Analyzed: 11:06

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	24790	17.211	16589	21.393	12046	23.712
	UPPER LIMIT	49580	17.711	33178	21.893	24092	24.212
	LOWER LIMIT	12395	16.711	8294.5	20.893	6023	23.212
	EPA SAMPLE NO.						
01	PB169903BL	16256	17.21	8725	21.39	8166	23.72
02	RW11-MW01S-20250922	23841	17.21	14883	21.39	11337	23.71
03	TT189D1-20250922	26628	17.21	18929	21.39	13815	23.71
04	RW11-MW01S-20250922MS	15556	17.21	10706	21.39	10137	23.71
05	RW11-MW01S-20250922MSD	15754	17.21	10676	21.39	10526	23.71
06	TT150S1-20250922	23771	17.21	14168	21.39	11452	23.71
07	RW8-MW01D2-20250922	24929	17.21	16069	21.39	11851	23.71
08	RW8-MW01D3-20250922	20950	17.21	13153	21.39	11624	23.71
09	PB169903BS	17535	17.21	9385	21.38	8566	23.71
10	TT163S1-20250924	24588	17.21	14333	21.38	11269	23.71
11	TT188S1-20250922	22848	17.21	13580	21.39	10518	23.71
12	TT192D2-20250922	22381	17.21	14037	21.39	10467	23.71
13	TT192D1-20250922	22305	17.21	12979	21.39	3373 *	23.71
14	DUP10-20250922	23623	17.21	14360	21.39	11948	23.71
15	RW11-MW01I-20250922	21770	17.21	13512	21.39	12343	23.71

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3194

Client ID: SSTDCCC0.4

Date Analyzed: 10/06/2025

Lab File ID: BN038004.D

Time Analyzed: 11:06

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	24790	17.211	16589	21.393	12046	23.712
UPPER LIMIT	49580	17.711	33178	21.893	24092	24.212
LOWER LIMIT	12395	16.711	8294.5	20.893	6023	23.212
EPA SAMPLE NO.						

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.