



Fax: 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Tab Name. ATTIANCE	CONCIACE. ALCOIS	
Lab Code: ACE	SDG NO.: Q3194	
Lab File ID: BN038021.D	DFTPP Injection Date:	10/06/2025
Instrument ID: BNA_N	DFTPP Injection Time:	22:13

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	40
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	15.3
442	Greater than 50% of mass 198	88.5
443	15.0 - 24.0% of mass 442	18.7 (21.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN038022.D	10/06/2025	23:32
RE139D1-20250924	Q3194-14	BN038024.D	10/07/2025	00:44
EB05-20250924	Q3194-15	BN038025.D	10/07/2025	01:21
EB06-20250924	Q3194-16	BN038026.D	10/07/2025	01:57
FB01-20250924	Q3194-17	BN038027.D	10/07/2025	02:33
SSTDCCC0.4EC	SSTDCCC0.4	BN038039.D	10/07/2025	09:47