



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100325\
Data File : VX047980.D
Acq On : 03 Oct 2025 08:35
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Title : SW846 8260
Last Update : Wed Sep 17 06:39:58 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1629

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.8	11857	PASS
75	95	30	60	53.1	30267	PASS
95	95	100	100	100.0	56984	PASS
96	95	5	9	6.5	3726	PASS
173	174	0.00	2	1.3	499	PASS
174	95	50	100	67.2	38299	PASS
175	174	5	9	7.7	2956	PASS
176	174	95	101	95.4	36533	PASS
177	176	5	9	6.4	2325	PASS