

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3194</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>AECOM Technical Services, Inc.</u>	Datafile :	<u>VX047928.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1001WBS01	Dichlorodifluoromethane	20	20.1	ug/L	101			32	152	
	Chloromethane	20	18.7	ug/L	94			50	139	
	Vinyl chloride	20	19.5	ug/L	98			58	137	
	Bromomethane	20	22.4	ug/L	112			53	141	
	Chloroethane	20	19.4	ug/L	97			60	138	
	Trichlorofluoromethane	20	19.5	ug/L	98			65	141	
	1,1,2-Trichlorotrifluoroethane	20	20.0	ug/L	100			70	136	
	1,1-Dichloroethene	20	19.5	ug/L	98			71	131	
	Acetone	100	95.2	ug/L	95			39	160	
	Carbon disulfide	20	18.2	ug/L	91			64	133	
	Methyl tert-butyl Ether	20	19.6	ug/L	98			71	124	
	Methyl Acetate	20	20.2	ug/L	101			56	136	
	Methylene Chloride	20	20.8	ug/L	104			74	124	
	trans-1,2-Dichloroethene	20	19.6	ug/L	98			75	124	
	1,1-Dichloroethane	20	20.6	ug/L	103			77	125	
	Cyclohexane	20	18.2	ug/L	91			71	130	
	2-Butanone	100	95.9	ug/L	96			56	143	
	Carbon Tetrachloride	20	19.8	ug/L	99			72	136	
	cis-1,2-Dichloroethene	20	20.7	ug/L	104			78	123	
	Bromochloromethane	20	19.4	ug/L	97			78	123	
	Chloroform	20	21.2	ug/L	106			79	124	
	1,1,1-Trichloroethane	20	20.3	ug/L	102			74	131	
	Methylcyclohexane	20	17.8	ug/L	89			72	132	
	Benzene	20	20.2	ug/L	101			79	120	
	1,2-Dichloroethane	20	20.4	ug/L	102			73	128	
	Trichloroethene	20	20.0	ug/L	100			79	123	
	1,2-Dichloropropane	20	21.1	ug/L	106			78	122	
	Bromodichloromethane	20	21.0	ug/L	105			79	125	
	4-Methyl-2-Pentanone	100	99.9	ug/L	100			67	130	
	Toluene	20	20.1	ug/L	101			80	121	
	t-1,3-Dichloropropene	20	19.9	ug/L	100			73	127	
	cis-1,3-Dichloropropene	20	20.2	ug/L	101			75	124	
	1,1,2-Trichloroethane	20	21.4	ug/L	107			80	119	
	2-Hexanone	100	96.8	ug/L	97			57	139	
	Dibromochloromethane	20	21.6	ug/L	108			74	126	
	1,2-Dibromoethane	20	21.3	ug/L	106			77	121	
	Tetrachloroethene	20	19.0	ug/L	95			74	129	
	Chlorobenzene	20	20.2	ug/L	101			82	118	
	Ethyl Benzene	20	19.3	ug/L	97			79	121	
	m/p-Xylenes	40	39.2	ug/L	98			80	121	
	o-Xylene	20	19.5	ug/L	98			78	122	
	Styrene	20	19.9	ug/L	100			78	123	
	Bromoform	20	20.3	ug/L	102			66	130	
	Isopropylbenzene	20	18.9	ug/L	95			72	131	
	1,1,2,2-Tetrachloroethane	20	19.7	ug/L	99			71	121	
	1,3-Dichlorobenzene	20	19.5	ug/L	98			80	119	
	1,4-Dichlorobenzene	20	19.4	ug/L	97			79	118	
	1,2-Dichlorobenzene	20	19.6	ug/L	98			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3194</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>AECOM Technical Services, Inc.</u>	Datafile :	<u>VX047928.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1001WBS01	1,2-Dibromo-3-Chloropropane	20	16.6	ug/L	83			62	128	
	1,2,4-Trichlorobenzene	20	18.1	ug/L	91			69	130	
	1,2,3-Trichlorobenzene	20	18.5	ug/L	93			69	129	