

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO15
 Lab Code: ACE SDG NO.: Q3194
 Lab File ID: VX047925.D Date Analyzed: 10/01/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:44
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	111244	12.00€				
UPPER LIMIT	222488	12.50€				
LOWER LIMIT	55622	11.50€				
EPA SAMPLE NO.						
TT189D1-20250922	139689	12.01				
RW8-MW01D2-20250922	127034	12.01				
RW8-MW01D3-20250922	127057	12.01				
TT188S1-20250922	124890	12.01				
VX1001WBL01	149007	12.01				
VX1001WBS01	117338	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.