

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

|                |                   |                |                   |
|----------------|-------------------|----------------|-------------------|
| Lab Name:      | <u>Alliance</u>   | Contract:      | <u>AECO15</u>     |
| Lab Code:      | <u>ACE</u>        | SDG NO.:       | <u>Q3194</u>      |
| Lab File ID:   | <u>VX047953.D</u> | Date Analyzed: | <u>10/02/2025</u> |
| Instrument ID: | <u>MSVOA_X</u>    | Time Analyzed: | <u>08:50</u>      |
| GC Column:     | <u>DB-624UI</u>   | ID:            | <u>0.18</u> (mm)  |
|                |                   | Heated Purge:  | (Y/N) <u>N</u>    |

|                        | IS4<br>AREA # | RT #  |  |  |  |  |
|------------------------|---------------|-------|--|--|--|--|
| 12 HOUR STD            | 107762        | 12    |  |  |  |  |
| UPPER LIMIT            | 215524        | 12.5  |  |  |  |  |
| LOWER LIMIT            | 53881         | 11.5  |  |  |  |  |
| EPA SAMPLE NO.         |               |       |  |  |  |  |
| RW11-MW01S-20250922    | 131301        | 12.01 |  |  |  |  |
| RW11-MW01S-20250922MS  | 117758        | 12.01 |  |  |  |  |
| RW11-MW01S-20250922MSD | 99094         | 12.01 |  |  |  |  |
| TT192D2-20250922       | 140515        | 12.01 |  |  |  |  |
| TT192D1-20250922       | 150104        | 12.01 |  |  |  |  |
| DUP10-20250922         | 129857        | 12.01 |  |  |  |  |
| RW11-MW01I-20250922    | 149635        | 12.01 |  |  |  |  |
| TT163S1-20250924       | 147498        | 12.01 |  |  |  |  |
| RE139D1-20250924       | 133075        | 12.01 |  |  |  |  |
| EB05-20250924          | 134089        | 12.01 |  |  |  |  |
| EB06-20250924          | 144409        | 12.01 |  |  |  |  |
| FB01-20250924          | 142205        | 12.01 |  |  |  |  |
| TB09-20250912          | 130557        | 12.01 |  |  |  |  |
| VX1002WBL01            | 160196        | 12.01 |  |  |  |  |
| VX1002WBS01            | 117908        | 12.01 |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.