

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TECH05
 Lab Code: ACE SDG No.: Q3195
 Instrument ID: MSVOA_V Calibration Date/Time: 09/26/2025 09:20
 Lab File ID: VV039198.D Init. Calib. Date(s): 09/22/2025 09/22/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 12:26 16:35
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Chloroethane	50.000	49.61	0.7720032	20
Acetone	250.000	325.48	-30.1904	20
Methylene Chloride	50.000	53.37	-6.746002	20
2-Hexanone	250.000	266.94	-6.776403	20
Styrene	50.000	50.58	-1.155998	20
1,2-Dibromo-3-Chloropropane	50.000	54.34	-8.678001	20

All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.