

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TECH05
 Lab Code: ACE SDG No.: Q3195
 Instrument ID: MSVOA_V Calibration Date/Time: 09/29/2025 11:11
 Lab File ID: VV039221.D Init. Calib. Date(s): 09/22/2025 09/22/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 12:26 16:35
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Chloroethane	50.000	45.57	8.862	20
Acetone	250.000	241.98	3.206799	20
Methylene Chloride	50.000	48.96	2.078003	20
2-Hexanone	250.000	255.23	-2.090002	20
Styrene	50.000	51.43	-2.863998	20
1,2-Dibromo-3-Chloropropane	50.000	52.56	-5.110001	20

All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.