

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TECH05
 Lab Code: ACE SDG NO.: Q3195
 Lab File ID: VV039168.D Date Analyzed: 09/25/2025
 Instrument ID: MSVOA_V Time Analyzed: 09:59
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	605297	4.60	933338	5.53	901987	8.77
UPPER LIMIT	1210590	5.096	1866680	6.029	1803970	9.273
LOWER LIMIT	302649	4.096	466669	5.029	450994	8.273
EPA SAMPLE NO.						
MW-1-091725	535790	4.60	971101	5.53	956100	8.78
MW-22-091825	525270	4.60	956838	5.54	943207	8.78
MW-23-092325	518728	4.60	934521	5.54	917520	8.78
MW-24-091925	515327	4.60	928629	5.54	925762	8.78
RW-1-092425	349429	4.60	628697	5.54	611801	8.78
EB-01-091925	360957	4.60	632244	5.54	624309	8.78
FB-01-091925	562300	4.60	1002078	5.54	977857	8.78
FB-02-092225	557526	4.60	1004933	5.53	984426	8.78
TB-01-091525	466916	4.60	828915	5.54	809498	8.78
VV0925WBL01	480047	4.60	864912	5.53	870930	8.78
VV0925WBS01	557017	4.60	898902	5.54	871912	8.78

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.