

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

|                |                   |                |                   |
|----------------|-------------------|----------------|-------------------|
| Lab Name:      | <u>Alliance</u>   | Contract:      | <u>TECH05</u>     |
| Lab Code:      | <u>ACE</u>        | SDG NO.:       | <u>Q3195</u>      |
| Lab File ID:   | <u>VV039198.D</u> | Date Analyzed: | <u>09/26/2025</u> |
| Instrument ID: | <u>MSVOA_V</u>    | Time Analyzed: | <u>09:20</u>      |
| GC Column:     | <u>DB-624UI</u>   | ID:            | <u>0.18</u> (mm)  |
|                |                   | Heated Purge:  | (Y/N) <u>N</u>    |

|                  | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #  |
|------------------|---------------|-------|---------------|-------|---------------|-------|
| 12 HOUR STD      | 568268        | 4.60  | 890856        | 5.53  | 851107        | 8.77  |
| UPPER LIMIT      | 1136540       | 5.097 | 1781710       | 6.029 | 1702210       | 9.273 |
| LOWER LIMIT      | 284134        | 4.097 | 445428        | 5.029 | 425554        | 8.273 |
| EPA SAMPLE NO.   |               |       |               |       |               |       |
| MOS-18-092225    | 335184        | 4.60  | 584917        | 5.54  | 609610        | 8.78  |
| MW-3-091525      | 599898        | 4.60  | 1062714       | 5.53  | 1062408       | 8.78  |
| MW-4-091625      | 513652        | 4.60  | 924164        | 5.54  | 911885        | 8.78  |
| MW-6-091625      | 374514        | 4.60  | 660135        | 5.54  | 675751        | 8.78  |
| MW-19D-092225    | 344932        | 4.60  | 614959        | 5.54  | 627934        | 8.78  |
| MW-21M-091925    | 513487        | 4.60  | 944501        | 5.54  | 957014        | 8.78  |
| MW-26-091925     | 486629        | 4.60  | 867427        | 5.54  | 873266        | 8.78  |
| MW-26-091925-FD  | 454191        | 4.60  | 813227        | 5.54  | 842320        | 8.78  |
| MW-29-092325     | 450640        | 4.60  | 819940        | 5.54  | 841715        | 8.78  |
| MW-31S-091825    | 444783        | 4.60  | 811223        | 5.54  | 838423        | 8.78  |
| MW-31D-091825    | 523421        | 4.60  | 967034        | 5.54  | 980974        | 8.78  |
| MW-31D-091825-FD | 333282        | 4.60  | 591720        | 5.54  | 612863        | 8.78  |
| EB-02-092225     | 509169        | 4.60  | 912297        | 5.54  | 888919        | 8.78  |
| VV0926WBL01      | 488972        | 4.60  | 870889        | 5.54  | 870759        | 8.78  |
| VV0926WBS01      | 545170        | 4.60  | 871681        | 5.53  | 834845        | 8.78  |

IS1 = Pentafluorobenzene  
IS2 = 1,4-Difluorobenzene  
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
AREA LOWER LIMIT = -50% of internal standard area  
RT UPPER LIMIT = +0.50 minutes of internal standard RT  
RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
\* Values outside of QC limits.