

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>TECH05</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3195</u>
Lab File ID:	<u>VV039168.D</u>	Date Analyzed:	<u>09/25/2025</u>
Instrument ID:	<u>MSVOA_V</u>	Time Analyzed:	<u>09:59</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	535710	11.172				
UPPER LIMIT	1071420	11.672				
LOWER LIMIT	267855	10.672				
EPA SAMPLE NO.						
MW-1-091725	458217	11.17				
MW-22-091825	451495	11.18				
MW-23-092325	439919	11.17				
MW-24-091925	442595	11.17				
RW-1-092425	288234	11.18				
EB-01-091925	293380	11.18				
FB-01-091925	478832	11.18				
FB-02-092225	478455	11.18				
TB-01-091525	384637	11.18				
VV0925WBL01	413592	11.18				
VV0925WBS01	508947	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.