

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TECH05
 Lab Code: ACE SDG NO.: Q3195
 Lab File ID: VV039198.D Date Analyzed: 09/26/2025
 Instrument ID: MSVOA_V Time Analyzed: 09:20
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	503461	11.172				
UPPER LIMIT	1006920	11.672				
LOWER LIMIT	251731	10.672				
EPA SAMPLE NO.						
MOS-18-092225	284053	11.18				
MW-3-091525	512995	11.18				
MW-4-091625	438479	11.17				
MW-6-091625	313063	11.18				
MW-19D-092225	291769	11.18				
MW-21M-091925	451121	11.18				
MW-26-091925	420754	11.18				
MW-26-091925-FD	398907	11.17				
MW-29-092325	391674	11.18				
MW-31S-091825	399971	11.17				
MW-31D-091825	469900	11.18				
MW-31D-091825-FD	282611	11.17				
EB-02-092225	432706	11.17				
VV0926WBL01	422120	11.18				
VV0926WBS01	490803	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.