

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>TECH05</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3195</u>
Lab File ID:	<u>VV039221.D</u>	Date Analyzed:	<u>09/29/2025</u>
Instrument ID:	<u>MSVOA_V</u>	Time Analyzed:	<u>11:11</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	505725	11.172				
UPPER LIMIT	1011450	11.672				
LOWER LIMIT	252863	10.672				
EPA SAMPLE NO.						
MW-19M-092325	378816	11.18				
MW-31S-091825DL	445110	11.18				
MW-31D-091825DL	405371	11.17				
MW-31D-091825-FDDL	404883	11.18				
MW-32S-091725	412882	11.18				
MW-33D-091725	398638	11.18				
MW-33D-091725MS	463779	11.17				
MW-33D-091725MSD	497049	11.17				
VV0929WBL01	403956	11.18				
VV0929WBS01	484149	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.