

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>M2AS01</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3201</u>
Instrument ID: <u>MSVOA_V</u>	Calibration Date(s): <u>09/22/2025</u> <u>09/22/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>12:26</u> <u>16:35</u>
GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm)	

LAB FILE ID: RRF010 = VV039137.D RRF020 = VV039138.D RRF050 = VV039139.D RRF100 = VV039140.D RRF005 = VV039142.D RRF001 = VV039143.D								
COMPOUND	RRF010	RRF020	RRF050	RRF100	RRF005	RRF001	RRF	% RSD
Dichlorodifluoromethane	0.848	0.757	0.740	0.699	0.772	0.934	0.792	10.8
Chloromethane	0.949	0.814	0.770	0.729	0.889	0.946	0.850	10.9
Vinyl Chloride	1.007	0.856	0.836	0.801	0.932	1.026	0.910	10.3
Bromomethane	0.631	0.548	0.508	0.524	0.655		0.573	11.5
Chloroethane	0.662	0.568	0.533	0.500	0.606	0.915	0.631	23.8
Trichlorofluoromethane	1.394	1.253	1.193	1.151	1.365	1.546	1.317	11.2
1,1,2-Trichlorotrifluoroethane	0.839	0.732	0.702	0.686	0.804	0.939	0.784	12.3
Tert butyl alcohol	0.169	0.157	0.137	0.136	0.166		0.153	10.3
1,1-Dichloroethene	0.804	0.721	0.674	0.663	0.789	0.873	0.754	10.9
Acetone	0.537	0.458	0.449	0.414	0.569	0.712	0.523	20.8
Carbon Disulfide	2.453	2.181	2.081	1.986	2.462	2.670	2.306	11.4
Methyl tert-butyl Ether	2.547	2.285	2.233	2.202	2.556	2.591	2.402	7.5
Methyl Acetate	1.146	1.036	0.983	0.973	1.097	1.198	1.072	8.5
Methylene Chloride	0.930	0.848	0.779	0.742	1.113	1.987	1.066	44
trans-1,2-Dichloroethene	0.846	0.755	0.724	0.696	0.835	0.959	0.803	12.1
1,1-Dichloroethane	1.669	1.466	1.388	1.329	1.659	1.942	1.576	14.4
Cyclohexane	1.325	1.222	1.245	1.254	1.407		1.291	5.9
2-Butanone	0.604	0.584	0.586	0.580	0.570	0.554	0.579	2.9
Carbon Tetrachloride	0.742	0.710	0.703	0.675	0.736	0.890	0.743	10.2
cis-1,2-Dichloroethene	0.908	0.842	0.868	0.860	0.832	1.001	0.885	7.1
Bromochloromethane	0.836	0.753	0.679	0.637	0.525	0.743	0.696	15.5
Chloroform	1.752	1.570	1.490	1.433	1.746	1.880	1.645	10.6
1,1,1-Trichloroethane	1.562	1.321	1.298	1.263	1.513	1.590	1.424	10.3
Methylcyclohexane	0.701	0.720	0.823	0.857	0.680	0.622	0.734	12.2
Benzene	2.077	1.978	1.990	1.931	1.963	1.936	1.979	2.7
1,2-Dichloroethane	0.725	0.694	0.666	0.643	0.693	0.740	0.693	5.2
Trichloroethene	0.471	0.456	0.463	0.454	0.460	0.435	0.457	2.6
1,2-Dichloropropane	0.509	0.519	0.502	0.486	0.471	0.465	0.492	4.4
Bromodichloromethane	0.832	0.764	0.753	0.725	0.767	0.829	0.778	5.5
4-Methyl-2-Pentanone	0.715	0.736	0.732	0.714	0.613	0.459	0.661	16.5

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: M2AS01

Lab Code: ACE

SDG No.: Q3201

Instrument ID: MSVOA_V

Calibration Date(s): 09/22/2025 09/22/2025

Heated Purge: (Y/N) N

Calibration Time(s): 12:26 16:35

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF010 = VV039137.D		RRF020 = VV039138.D		RRF050 = VV039139.D		
		RRF100 = VV039140.D		RRF005 = VV039142.D		RRF001 = VV039143.D		
COMPOUND	RRF010	RRF020	RRF050	RRF100	RRF005	RRF001	RRF	% RSD
Toluene	1.211	1.199	1.229	1.214	1.130	0.901	1.147	10.9
t-1,3-Dichloropropene	0.701	0.696	0.728	0.753	0.645	0.548	0.678	10.8
cis-1,3-Dichloropropene	0.803	0.777	0.811	0.803	0.704	0.595	0.749	11.4
1,1,2-Trichloroethane	0.534	0.502	0.494	0.480	0.524	0.553	0.514	5.3
2-Hexanone	0.505	0.534	0.546	0.533	0.429	0.231	0.463	26.2
Dibromochloromethane	0.605	0.570	0.567	0.558	0.598	0.620	0.586	4.2
1,2-Dibromoethane	0.556	0.538	0.529	0.514	0.535	0.484	0.526	4.7
Tetrachloroethene	0.435	0.412	0.412	0.407	0.431	0.415	0.419	2.7
Chlorobenzene	1.403	1.368	1.383	1.358	1.417	1.428	1.393	2
Ethyl Benzene	2.028	2.143	2.382	2.418	1.953	1.725	2.108	12.5
m/p-Xylenes	0.820	0.882	0.949	0.938	0.713	0.580	0.814	17.7
o-Xylene	0.703	0.742	0.846	0.880	0.635	0.518	0.720	18.7
Styrene	1.239	1.381	1.553	1.537	1.038	0.828	1.263	22.8
Bromoform	0.396	0.384	0.396	0.400	0.385	0.397	0.393	1.7
Isopropylbenzene	3.564	3.638	3.939	4.096	3.296	3.226	3.627	9.5
1,1,2,2-Tetrachloroethane	1.600	1.461	1.425	1.389	1.684	1.975	1.589	13.8
1,3-Dichlorobenzene	1.931	1.836	1.857	1.882	1.943	1.975	1.904	2.8
1,4-Dichlorobenzene	2.070	1.951	1.934	1.919	2.139	2.522	2.089	11
1,2-Dichlorobenzene	1.751	1.652	1.738	1.802	1.797	1.935	1.779	5.3
1,2-Dibromo-3-Chloropropane	0.299	0.283	0.285	0.302	0.333	0.476	0.330	22.4
1,2,4-Trichlorobenzene	0.928	0.963	1.076	1.189	0.935	1.028	1.020	9.9
1,2,3-Trichlorobenzene	0.974	0.973	1.107	1.191	0.901	1.113	1.043	10.6
1,2-Dichloroethane-d4	0.773	0.663	0.646	0.618	0.919		0.724	17.1
Dibromofluoromethane	0.435	0.382	0.377	0.360	0.454		0.402	10.1
Toluene-d8	1.166	1.092	1.136	1.097	1.411		1.181	11.2
4-Bromofluorobenzene	0.481	0.464	0.507	0.501	0.477		0.486	3.6

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.