

Instrument ID: **MSVOA_V**

Daily Analysis Runlog For Sequence/QCBatch ID # VV092225

Review By	Mahesh Dadoda	Review On	9/26/2025 1:39:32 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	9/26/2025 1:56:56 AM		
SubDirectory	VV092225	HP Acquire Method	MSVOA_V	HP Processing Method	82v092225w.m
STD. NAME		STD REF.#			
Tune/Reschk		VP135580			
Initial Calibration Stds		VP135581,VP135582,VP135583,VP135584,VP135585,VP135586			
CCC					
Internal Standard/PEM		VP133935			
ICV/I.BLK		VP135587			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VV039134.D	22 Sep 2025 10:15	SY/MD	Ok
2	VSTDIC001	VV039135.D	22 Sep 2025 10:59	SY/MD	Not Ok
3	VSTDIC005	VV039136.D	22 Sep 2025 11:30	SY/MD	Not Ok
4	VSTDIC010	VV039137.D	22 Sep 2025 12:26	SY/MD	Ok
5	VSTDIC020	VV039138.D	22 Sep 2025 12:48	SY/MD	Ok
6	VSTDIC050	VV039139.D	22 Sep 2025 13:26	SY/MD	Ok
7	VSTDIC100	VV039140.D	22 Sep 2025 13:49	SY/MD	Ok
8	VIBLK	VV039141.D	22 Sep 2025 14:21	SY/MD	Ok
9	VSTDIC005	VV039142.D	22 Sep 2025 15:37	SY/MD	Ok,M
10	VSTDIC001	VV039143.D	22 Sep 2025 16:35	SY/MD	Ok,M
11	VSTDICV050	VV039144.D	22 Sep 2025 16:58	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_V**

Daily Analysis Runlog For Sequence/QC Batch ID # VV100325

Review By	Mahesh Dadoda	Review On	10/6/2025 4:25:08 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	10/6/2025 4:29:07 AM		
SubDirectory	VV100325	HP Acquire Method	MSVOA_V	HP Processing Method	82v092225w.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP135740			
CCC		VP135741,VP135742			
Internal Standard/PEM		VP133935			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VV039241.D	03 Oct 2025 08:26	SY/MD	Ok
2	VSTDCCC050	VV039242.D	03 Oct 2025 08:53	SY/MD	Ok
3	VV1003MBL01	VV039243.D	03 Oct 2025 11:04	SY/MD	Ok
4	VV1003WBL01	VV039244.D	03 Oct 2025 11:26	SY/MD	Ok
5	VV1003WBS01	VV039245.D	03 Oct 2025 13:07	SY/MD	Ok
6	VV1003WBSD01	VV039246.D	03 Oct 2025 13:29	SY/MD	Ok
7	Q3263-01	VV039247.D	03 Oct 2025 14:19	SY/MD	ReRun
8	Q3263-02	VV039248.D	03 Oct 2025 14:42	SY/MD	ReRun
9	Q3201-07	VV039249.D	03 Oct 2025 15:04	SY/MD	Ok
10	Q3201-08	VV039250.D	03 Oct 2025 15:27	SY/MD	Ok
11	Q3254-09	VV039251.D	03 Oct 2025 15:49	SY/MD	Ok
12	Q3254-01	VV039252.D	03 Oct 2025 16:11	SY/MD	Ok
13	Q3254-03	VV039253.D	03 Oct 2025 16:34	SY/MD	Ok
14	Q3254-05	VV039254.D	03 Oct 2025 16:56	SY/MD	Ok
15	Q3254-07	VV039255.D	03 Oct 2025 17:19	SY/MD	Ok
16	Q3201-01	VV039256.D	03 Oct 2025 17:41	SY/MD	Dilution
17	Q3201-02	VV039257.D	03 Oct 2025 18:04	SY/MD	Dilution
18	Q3201-03	VV039258.D	03 Oct 2025 18:26	SY/MD	ReRun
19	Q3201-04	VV039259.D	03 Oct 2025 18:49	SY/MD	Not Ok
20	Q3201-05	VV039260.D	03 Oct 2025 19:11	SY/MD	Dilution
21	Q3201-06	VV039261.D	03 Oct 2025 19:34	SY/MD	Not Ok

Instrument ID: **MSVOA_V**

Daily Analysis Runlog For Sequence/QCBatch ID # VV100325

Review By	Mahesh Dadoda	Review On	10/6/2025 4:25:08 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	10/6/2025 4:29:07 AM		
SubDirectory	VV100325	HP Acquire Method	MSVOA_V	HP Processing Method	82v092225w.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP135740			
CCC		VP135741,VP135742			
Internal Standard/PEM		VP133935			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	VIBLK	VV039262.D	03 Oct 2025 19:56	SY/MD	Ok
23	VSTDCCC050	VV039263.D	03 Oct 2025 20:19	SY/MD	Ok

M : Manual Integration

Instrument ID: **MSVOA_V**

Daily Analysis Runlog For Sequence/QC Batch ID # VV100625

Review By	Mahesh Dadoda	Review On	10/7/2025 1:40:31 PM
Supervise By	Semsettin Yesilyurt	Supervise On	10/7/2025 1:42:17 PM
SubDirectory	VV100625	HP Acquire Method	MSVOA_V
		HP Processing Method	
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135749		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135750,VP135751 VP133935		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VV039264.D	06 Oct 2025 08:07	SY/MD	Ok
2	VSTDCCC050	VV039265.D	06 Oct 2025 11:46	SY/MD	Ok
3	VV1006MBL01	VV039266.D	06 Oct 2025 12:13	SY/MD	Ok
4	VV1006WBL01	VV039267.D	06 Oct 2025 12:35	SY/MD	Ok
5	VV1006WBS01	VV039268.D	06 Oct 2025 13:01	SY/MD	Ok
6	VV1006WBSD01	VV039269.D	06 Oct 2025 13:24	SY/MD	Ok
7	Q3201-05DL	VV039270.D	06 Oct 2025 13:56	SY/MD	Ok,M
8	Q3201-01DL	VV039271.D	06 Oct 2025 14:19	SY/MD	Ok
9	Q3201-02DL	VV039272.D	06 Oct 2025 14:41	SY/MD	Ok,M
10	PB169972TB	VV039273.D	06 Oct 2025 15:03	SY/MD	Ok,M
11	Q3269-03	VV039274.D	06 Oct 2025 15:26	SY/MD	Ok
12	Q3269-06	VV039275.D	06 Oct 2025 15:48	SY/MD	Ok
13	Q3289-02	VV039276.D	06 Oct 2025 16:11	SY/MD	ReRun
14	Q3289-04	VV039277.D	06 Oct 2025 16:33	SY/MD	ReRun
15	Q3289-06	VV039278.D	06 Oct 2025 16:56	SY/MD	Ok
16	VIBLK	VV039279.D	06 Oct 2025 17:18	SY/MD	Ok
17	VSTDCCC050	VV039280.D	06 Oct 2025 17:40	SY/MD	Ok

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX091625

Review By	John Carlone	Review On	9/17/2025 8:27:00 AM
Supervise By	Maresh Dadoda	Supervise On	9/18/2025 3:22:52 PM
SubDirectory	VX091625	HP Acquire Method	HP Processing Method 82X091625W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP135487		
Initial Calibration Stds	VP135489,VP135490,VP135491,VP135492,VP135493,VP135494		
CCC	VP135488		
Internal Standard/PEM			
ICV/I.BLK	VP135495		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX047584.D	16 Sep 2025 08:56	JC/MD	Ok
2	VSTDIC001	VX047585.D	16 Sep 2025 09:23	JC/MD	Ok,M
3	VSTDIC005	VX047586.D	16 Sep 2025 10:16	JC/MD	Ok,M
4	VSTDIC020	VX047587.D	16 Sep 2025 10:37	JC/MD	Ok,M
5	VSTDIC050	VX047588.D	16 Sep 2025 10:58	JC/MD	Ok,M
6	VSTDIC100	VX047589.D	16 Sep 2025 11:40	JC/MD	Ok,M
7	VSTDIC150	VX047590.D	16 Sep 2025 12:01	JC/MD	Ok,M
8	IBLK	VX047591.D	16 Sep 2025 12:22	JC/MD	Ok
9	VSTDICV050	VX047592.D	16 Sep 2025 13:35	JC/MD	Ok,M
10	VX0916MBL01	VX047593.D	16 Sep 2025 14:03	JC/MD	Ok
11	VX0916WBL01	VX047594.D	16 Sep 2025 14:53	JC/MD	Ok
12	VX0916WBS01	VX047595.D	16 Sep 2025 15:21	JC/MD	Ok,M
13	VX0916WBSD01	VX047596.D	16 Sep 2025 15:46	JC/MD	Ok,M
14	Q3015-01	VX047597.D	16 Sep 2025 16:07	JC/MD	Ok,M
15	VIBLK	VX047598.D	16 Sep 2025 16:30	JC/MD	Ok
16	VSTDIC050	VX047599.D	16 Sep 2025 16:51	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX100625

Review By	John Carlone	Review On	10/7/2025 9:29:22 AM
Supervise By	Maresh Dadoda	Supervise On	10/8/2025 2:23:55 AM
SubDirectory	VX100625	HP Acquire Method	HP Processing Method 82X091625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135743		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135744,VP135745		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX048008.D	06 Oct 2025 08:12	JC/MD	Ok
2	VSTDCCC050	VX048009.D	06 Oct 2025 08:54	JC/MD	Ok,M
3	VX1006MBL01	VX048010.D	06 Oct 2025 09:23	JC/MD	Ok
4	VX1006WBL01	VX048011.D	06 Oct 2025 09:44	JC/MD	Ok
5	VX1006WBS01	VX048012.D	06 Oct 2025 10:44	JC/MD	Not Ok
6	Q3202-19	VX048013.D	06 Oct 2025 11:13	JC/MD	Ok
7	VX1006WBS02	VX048014.D	06 Oct 2025 11:34	JC/MD	Ok,M
8	VX1006WBSD02	VX048015.D	06 Oct 2025 12:01	JC/MD	Ok,M
9	Q3279-01	VX048016.D	06 Oct 2025 12:22	JC/MD	Ok
10	Q3279-02	VX048017.D	06 Oct 2025 12:43	JC/MD	Ok
11	Q3279-03	VX048018.D	06 Oct 2025 13:04	JC/MD	Ok
12	Q3279-04	VX048019.D	06 Oct 2025 13:25	JC/MD	Ok
13	Q3201-06	VX048020.D	06 Oct 2025 13:46	JC/MD	Ok,M
14	Q3201-03	VX048021.D	06 Oct 2025 14:07	JC/MD	Ok
15	Q3201-04	VX048022.D	06 Oct 2025 14:28	JC/MD	Ok
16	Q3263-01	VX048023.D	06 Oct 2025 14:49	JC/MD	Ok
17	Q3263-02	VX048024.D	06 Oct 2025 15:10	JC/MD	Ok
18	Q3202-03	VX048025.D	06 Oct 2025 15:31	JC/MD	Ok
19	Q3202-05	VX048026.D	06 Oct 2025 15:51	JC/MD	Ok
20	Q3202-07	VX048027.D	06 Oct 2025 16:12	JC/MD	Ok
21	Q3202-09	VX048028.D	06 Oct 2025 16:33	JC/MD	Ok

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX100625

Review By	John Carlone	Review On	10/7/2025 9:29:22 AM
Supervise By	Maresh Dadoda	Supervise On	10/8/2025 2:23:55 AM
SubDirectory	VX100625	HP Acquire Method	HP Processing Method 82X091625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135743		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135744,VP135745		

22	Q3202-17	VX048029.D	06 Oct 2025 16:54	JC/MD	Ok
23	IBLK	VX048030.D	06 Oct 2025 17:15	JC/MD	Ok
24	Q3202-11	VX048031.D	06 Oct 2025 17:35	JC/MD	Ok
25	Q3202-13MS	VX048032.D	06 Oct 2025 17:56	JC/MD	Ok,M
26	Q3202-15MSD	VX048033.D	06 Oct 2025 18:17	JC/MD	Ok,M
27	VSTDCCC050	VX048034.D	06 Oct 2025 18:38	JC/MD	Ok,M

M : Manual Integration