

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: M2AS01
 Lab Code: ACE SDG NO.: Q3201
 Lab File ID: VV039265.D Date Analyzed: 10/06/2025
 Instrument ID: MSVOA_V Time Analyzed: 11:46
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	549792	4.60	857645	5.53	835774	8.77
UPPER LIMIT	1099580	5.096	1715290	6.029	1671550	9.273
LOWER LIMIT	274896	4.096	428823	5.029	417887	8.273
EPA SAMPLE NO.						
MW6DL	477065	4.60	810750	5.53	821368	8.78
MW14DL	444773	4.60	756536	5.54	773912	8.78
MW1DL	525555	4.60	878942	5.53	886561	8.77
VV1006WBL01	463806	4.60	783914	5.54	804136	8.78
VV1006WBS01	527068	4.60	830218	5.53	815170	8.78

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.