

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>M2AS01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3201</u>
Lab File ID:	<u>VV039242.D</u>	Date Analyzed:	<u>10/03/2025</u>
Instrument ID:	<u>MSVOA_V</u>	Time Analyzed:	<u>08:53</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	467640	11.172				
UPPER LIMIT	935280	11.672				
LOWER LIMIT	233820	10.672				
EPA SAMPLE NO.						
MW6	285289	11.18				
MW14	341647	11.17				
MW1	431237	11.18				
FIELD-BLANK	300076	11.18				
TRIP-BLANK	264732	11.18				
VV1003WBL01	433083	11.17				
VV1003WBS01	506399	11.17				
VV1003WBSD01	481201	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.