

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>M2AS01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3201</u>
Lab File ID:	<u>VV039265.D</u>	Date Analyzed:	<u>10/06/2025</u>
Instrument ID:	<u>MSVOA_V</u>	Time Analyzed:	<u>11:46</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	504333	11.172				
UPPER LIMIT	1008670	11.672				
LOWER LIMIT	252167	10.672				
EPA SAMPLE NO.						
MW6DL	441845	11.18				
MW14DL	417747	11.18				
MW1DL	514429	11.17				
VV1006WBL01	428242	11.17				
VV1006WBS01	499953	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.