

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>M2AS01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3201</u>
Lab File ID:	<u>VX048009.D</u>	Date Analyzed:	<u>10/06/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>08:54</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	117957	12				
UPPER LIMIT	235914	12.5				
LOWER LIMIT	58978.5	11.5				
EPA SAMPLE NO.						
MW15	149616	12.01				
MW10	151792	12.01				
MW3	159943	12.01				
VX1006WBL01	171990	12.01				
VX1006WBS02	126701	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.