

Analytical Summary Report

Analysis Method: SM 4500-H B
Parameter: pH
Run Number: LB137324

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/26/2025	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	09/26/2025	08:01
3	CAL3	1	Water	NA	NA	20.2	10.02	09/26/2025	08:05
4	ICV	1	Water	NA	NA	20.3	7.00	09/26/2025	08:10
5	CCV1	1	Water	NA	NA	20.3	2.01	09/26/2025	08:11
6	Q3202-01	1	Water	NA	NA	20.1	5.99	09/26/2025	08:15
7	Q3202-03	1	Water	NA	NA	20.6	5.46	09/26/2025	08:22
8	Q3202-05	1	Water	NA	NA	20.1	5.84	09/26/2025	08:30
9	Q3202-07	1	Water	NA	NA	20.1	5.94	09/26/2025	08:37
10	Q3202-09	1	Water	NA	NA	20.1	5.84	09/26/2025	08:45
11	Q3202-09DUP	1	Water	NA	NA	20.3	5.86	09/26/2025	08:47
12	Q3202-11	1	Water	NA	NA	20.1	6.44	09/26/2025	08:55
13	Q3202-17	1	Water	NA	NA	20.1	6.23	09/26/2025	09:00
14	CCV2	1	Water	NA	NA	20.3	12.02	09/26/2025	09:05

WORKLIST(Hardcopy Internal Chain)

WB 134324

WorkList Name : ph nq3204 w

WorkList ID : 192103

Department : Wet-Chemistry

Date : 09-26-2025 07:44:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3202-01	SY-6	Water	pH	Cool 4 deg C	LOCK01	I63	09/23/2025	SM 4500-H B
Q3202-03	SY-2R	Water	pH	Cool 4 deg C	LOCK01	I63	09/23/2025	SM 4500-H B
Q3202-05	SY-2D	Water	pH	Cool 4 deg C	LOCK01	I63	09/23/2025	SM 4500-H B
Q3202-07	SY-5	Water	pH	Cool 4 deg C	LOCK01	I63	09/23/2025	SM 4500-H B
Q3202-09	SY-3DD	Water	pH	Cool 4 deg C	LOCK01	I63	09/23/2025	SM 4500-H B
Q3202-11	SY-3D	Water	pH	Cool 4 deg C	LOCK01	I63	09/23/2025	SM 4500-H B
Q3202-17	SY-3	Water	pH	Cool 4 deg C	LOCK01	I63	09/23/2025	SM 4500-H B

Date/Time 09/26/25 07:45

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 09/26/25 13:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]