



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3202

Contract: LOCK01

Lab Code: ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Antimony	50.1	50.0	100	65 - 135	P	10/01/2025	12:49	LB137389
	Arsenic	21.9	20.0	110	65 - 135	P	10/01/2025	12:49	LB137389
	Barium	97.8	100	98	65 - 135	P	10/01/2025	12:49	LB137389
	Beryllium	6.36	6.0	106	65 - 135	P	10/01/2025	12:49	LB137389
	Cadmium	5.93	6.0	99	65 - 135	P	10/01/2025	12:49	LB137389
	Chromium	10.4	10.0	104	65 - 135	P	10/01/2025	12:49	LB137389
	Cobalt	29.6	30.0	99	65 - 135	P	10/01/2025	12:49	LB137389
	Copper	21.3	20.0	106	65 - 135	P	10/01/2025	12:49	LB137389
	Iron	103	100	103	65 - 135	P	10/01/2025	12:49	LB137389
	Lead	11.5	12.0	96	65 - 135	P	10/01/2025	12:49	LB137389
	Manganese	21.7	20.0	109	65 - 135	P	10/01/2025	12:49	LB137389
	Nickel	39.8	40.0	100	65 - 135	P	10/01/2025	12:49	LB137389
	Potassium	1910	2000	96	65 - 135	P	10/01/2025	12:49	LB137389
	Selenium	19.0	20.0	95	65 - 135	P	10/01/2025	12:49	LB137389
	Silver	10.1	10.0	101	65 - 135	P	10/01/2025	12:49	LB137389
	Sodium	1930	2000	96	65 - 135	P	10/01/2025	12:49	LB137389
	Thallium	42.2	40.0	105	65 - 135	P	10/01/2025	12:49	LB137389
	Vanadium	40.6	40.0	102	65 - 135	P	10/01/2025	12:49	LB137389
	Zinc	41.8	40.0	105	65 - 135	P	10/01/2025	12:49	LB137389