



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3202

Contract: LOCK01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	12:45	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	12:45	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	12:45	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	12:45	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	12:45	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	12:45	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	12:45	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	12:45	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	12:45	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	12:45	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	12:45	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	12:45	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	12:45	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	12:45	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	12:45	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	12:45	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	12:45	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	12:45	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	12:45	LB137389

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3202

Contract: LOCK01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	13:14	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	13:14	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	13:14	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	13:14	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	13:14	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	13:14	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	13:14	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	13:14	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	13:14	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	13:14	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	13:14	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	13:14	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	13:14	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	13:14	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	13:14	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	13:14	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	13:14	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	13:14	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	13:14	LB137389
CCB02	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	14:06	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	14:06	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	14:06	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	14:06	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	14:06	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	14:06	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	14:06	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	14:06	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	14:06	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	14:06	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	14:06	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	14:06	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	14:06	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	14:06	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	14:06	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	14:06	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	14:06	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	14:06	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	14:06	LB137389
CCB03	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	15:04	LB137389

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3202

Contract: LOCK01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	15:04	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	15:04	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	15:04	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	15:04	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	15:04	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	15:04	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	15:04	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	15:04	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	15:04	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	15:04	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	15:04	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	15:04	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	15:04	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	15:04	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	15:04	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	15:04	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	15:04	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	15:04	LB137389
CCB04	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	15:55	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	15:55	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	15:55	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	15:55	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	15:55	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	15:55	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	15:55	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	15:55	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	15:55	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	15:55	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	15:55	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	15:55	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	15:55	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	15:55	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	15:55	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	15:55	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	15:55	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	15:55	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	15:55	LB137389
CCB05	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	16:46	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	16:46	LB137389

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3202

Contract: LOCK01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Barium	14.6	+/-50	U	100	P	10/01/2025	16:46	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	16:46	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	16:46	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	16:46	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	16:46	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	16:46	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	16:46	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	16:46	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	16:46	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	16:46	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	16:46	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	16:46	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	16:46	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	16:46	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	16:46	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	16:46	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	16:46	LB137389
CCB06	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	17:41	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	17:41	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	17:41	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	17:41	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	17:41	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	17:41	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	17:41	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	17:41	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	17:41	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	17:41	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	17:41	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	17:41	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	17:41	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	17:41	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	17:41	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	17:41	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	17:41	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	17:41	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	17:41	LB137389
CCB07	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	18:32	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	18:32	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	18:32	LB137389

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3202

Contract: LOCK01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB07	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	18:32	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	18:32	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	18:32	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	18:32	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	18:32	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	18:32	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	18:32	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	18:32	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	18:32	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	18:32	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	18:32	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	18:32	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	18:32	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	18:32	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	18:32	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	18:32	LB137389
CCB08	Antimony	6.76	+/-25	U	50.0	P	10/01/2025	19:23	LB137389
	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	19:23	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	19:23	LB137389
	Beryllium	0.56	+/-3	U	6.00	P	10/01/2025	19:23	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	19:23	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	19:23	LB137389
	Cobalt	2.26	+/-15	U	30.0	P	10/01/2025	19:23	LB137389
	Copper	4.60	+/-10	U	20.0	P	10/01/2025	19:23	LB137389
	Iron	23.4	+/-50	U	100	P	10/01/2025	19:23	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	19:23	LB137389
	Manganese	5.94	+/-10	U	20.0	P	10/01/2025	19:23	LB137389
	Nickel	3.06	+/-20	U	40.0	P	10/01/2025	19:23	LB137389
	Potassium	918	+/-1000	U	2000	P	10/01/2025	19:23	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	19:23	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	19:23	LB137389
	Sodium	868	+/-1000	U	2000	P	10/01/2025	19:23	LB137389
	Thallium	4.38	+/-20	U	40.0	P	10/01/2025	19:23	LB137389
	Vanadium	6.26	+/-20	U	40.0	P	10/01/2025	19:23	LB137389
	Zinc	16.7	+/-20	U	40.0	P	10/01/2025	19:23	LB137389