

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3202

Contract: LOCK01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169863BS							
Antimony	ug/L	400	395		99	80 - 120	P
Arsenic	ug/L	400	380		95	80 - 120	P
Barium	ug/L	100	95.8		96	80 - 120	P
Beryllium	ug/L	100	97.1		97	80 - 120	P
Cadmium	ug/L	100	91.3		91	80 - 120	P
Chromium	ug/L	200	195		98	80 - 120	P
Cobalt	ug/L	100	93.9		94	80 - 120	P
Copper	ug/L	150	148		99	80 - 120	P
Iron	ug/L	1500	1410		94	80 - 120	P
Lead	ug/L	500	459		92	80 - 120	P
Manganese	ug/L	100	101		101	80 - 120	P
Nickel	ug/L	250	235		94	80 - 120	P
Potassium	ug/L	5000	4400		88	80 - 120	P
Selenium	ug/L	1000	940		94	80 - 120	P
Silver	ug/L	37.5	34.9		93	80 - 120	P
Sodium	ug/L	1500	1260		84	80 - 120	P
Thallium	ug/L	1000	885		88	80 - 120	P
Vanadium	ug/L	150	146		97	80 - 120	P
Zinc	ug/L	100	97.9		98	80 - 120	P