

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: LOCK01  
 Lab Code: ACE SDG NO.: Q3202  
 Lab File ID: VX047981.D Date Analyzed: 10/03/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 09:35  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 105470        | 12.00€ |  |  |  |  |
| UPPER LIMIT    | 210940        | 12.50€ |  |  |  |  |
| LOWER LIMIT    | 52735         | 11.50€ |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| SY-6           | 147037        | 12.01  |  |  |  |  |
| VX1003WBL01    | 147497        | 12.01  |  |  |  |  |
| VX1003WBS02    | 101886        | 12.01  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.