

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: LOCK01
 Lab Code: ACE SDG NO.: Q3202
 Lab File ID: VX048009.D Date Analyzed: 10/06/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:54
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	117957	12				
UPPER LIMIT	235914	12.5				
LOWER LIMIT	58978.5	11.5				
EPA SAMPLE NO.						
SY-2R	135085	12.01				
SY-2D	145340	12.01				
SY-5	131834	12.01				
SY-3DD	140531	12.01				
SY-3D	139753	12.01				
SY-3DMS	111123	12.01				
SY-3DMSD	95020	12.01				
SY-3	114487	12.01				
TB	161049	12.01				
VX1006WBL01	171990	12.01				
VX1006WBS02	126701	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.