



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Hit Summary Sheet**  
**SW-846**

<b>SDG No.:</b>	Q3207	<b>Order ID:</b>	Q3207
<b>Client:</b>	CHA Companies, Inc.	<b>Project ID:</b>	NYSOMH Emergency Service

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
<b>Client ID :</b> Q3207-01	<b>Drum#1- Poly Sheeting</b> Drum#1- Poly Sheeting	TCLP	Lead	583		11.5	60.0	ug/L
<b>Client ID :</b> Q3207-02	<b>Drum#2 PPE</b> Drum#2 PPE	TCLP	Lead	55500		11.5	60.0	ug/L
<b>Client ID :</b> Q3207-03	<b>Drum#3 Paint Chips</b> Drum#3 Paint Chips	TCLP	Lead	2600		11.5	60.0	ug/L



# SAMPLE DATA

A

B

C

D

## Report of Analysis

Client:	CHA Companies, Inc.	Date Collected:	09/25/25
Project:	NYSOMH Emergency Service	Date Received:	09/25/25
Client Sample ID:	Drum#1- Poly Sheeting	SDG No.:	Q3207
Lab Sample ID:	Q3207-01	Matrix:	TCLP
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-92-1	Lead	583		1	11.5	60.0	ug/L	09/30/25 12:30	09/30/25 21:04	6010D	SW3050

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	TCLP Metals Group3			

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N = Spiked sample recovery not within control limits

## Report of Analysis

Client:	CHA Companies, Inc.	Date Collected:	09/25/25
Project:	NYSOMH Emergency Service	Date Received:	09/25/25
Client Sample ID:	Drum#2 PPE	SDG No.:	Q3207
Lab Sample ID:	Q3207-02	Matrix:	TCLP
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-92-1	Lead	55500	1	11.5		60.0	ug/L	09/30/25 12:30	09/30/25 21:08	6010D	SW3050

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	TCLP Metals Group3			

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N = Spiked sample recovery not within control limits

## Report of Analysis

Client:	CHA Companies, Inc.	Date Collected:	09/25/25
Project:	NYSOMH Emergency Service	Date Received:	09/25/25
Client Sample ID:	Drum#3 Paint Chips	SDG No.:	Q3207
Lab Sample ID:	Q3207-03	Matrix:	TCLP
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-92-1	Lead	2600	1	11.5		60.0	ug/L	09/30/25 12:30	09/30/25 21:12	6010D	SW3050

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	TCLP Metals Group3			

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N = Spiked sample recovery not within control limits

## LAB CHRONICLE

<b>OrderID:</b> Q3207	<b>OrderDate:</b> 9/25/2025 12:17:00 PM
<b>Client:</b> CHA Companies, Inc.	<b>Project:</b> NYSOMH Emergency Service
<b>Contact:</b> Carrie Robinson	<b>Location:</b> J12

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3207-01	Drum#1- Poly Sheeting	TCLP			09/25/25	09/30/25	09/30/25	09/25/25
			TCLPMetals Group3	6010D				
Q3207-02	Drum#2 PPE	TCLP			09/25/25	09/30/25	09/30/25	09/25/25
			TCLPMetals Group3	6010D				
Q3207-03	Drum#3 Paint Chips	TCLP			09/25/25	09/30/25	09/30/25	09/25/25
			TCLPMetals Group3	6010D				