

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CHA Companies, Inc.

SDG No.: Q3207

Contract: CLOU03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3830	4000	96	90 - 110	P	09/30/2025	13:16	LB137374

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CHA Companies, Inc.

SDG No.: Q3207

Contract: CLOU03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.0	12.0	91	80 - 120	P	09/30/2025	13:22	LB137374

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CHA Companies, Inc.

SDG No.: Q3207

Contract: CLOU03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4850	5000	97	90 - 110	P	09/30/2025	14:08	LB137374
CCV02	Lead	4890	5000	98	90 - 110	P	09/30/2025	15:41	LB137374
CCV03	Lead	4890	5000	98	90 - 110	P	09/30/2025	16:43	LB137374
CCV04	Lead	4870	5000	98	90 - 110	P	09/30/2025	17:39	LB137374
CCV05	Lead	4840	5000	97	90 - 110	P	09/30/2025	18:55	LB137374
CCV06	Lead	4930	5000	99	90 - 110	P	09/30/2025	19:45	LB137374
CCV07	Lead	4820	5000	96	90 - 110	P	09/30/2025	20:35	LB137374
CCV08	Lead	4950	5000	99	90 - 110	P	09/30/2025	21:25	LB137374
CCV09	Lead	4860	5000	97	90 - 110	P	09/30/2025	22:14	LB137374
CCV10	Lead	5000	5000	100	90 - 110	P	09/30/2025	23:00	LB137374

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CHA Companies, Inc.

SDG No.: Q3207

Contract: CLOU03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3860	4000	96	90 - 110	P	10/01/2025	12:33	LB137389



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CHA Companies, Inc.

SDG No.: Q3207

Contract: CLOU03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.1	12.0	93	80 - 120	P	10/01/2025	12:41	LB137389

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CHA Companies, Inc.

SDG No.: Q3207

Contract: CLOU03

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4950	5000	99	90 - 110	P	10/01/2025	13:10	LB137389
CCV02	Lead	4960	5000	99	90 - 110	P	10/01/2025	14:02	LB137389
CCV03	Lead	5050	5000	101	90 - 110	P	10/01/2025	15:00	LB137389
CCV04	Lead	5000	5000	100	90 - 110	P	10/01/2025	15:51	LB137389
CCV05	Lead	4960	5000	99	90 - 110	P	10/01/2025	16:42	LB137389
CCV06	Lead	5030	5000	100	90 - 110	P	10/01/2025	17:32	LB137389
CCV07	Lead	4970	5000	99	90 - 110	P	10/01/2025	18:28	LB137389
CCV08	Lead	4860	5000	97	90 - 110	P	10/01/2025	19:19	LB137389