

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: CHA Companies, Inc.

SDG No.: Q3207

Contract: CLOU03

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	-6.58			-12	12	09/30/2025	13:44	LB137374
ICSAB01	Lead	38.9	49.0	79	37	61	09/30/2025	13:56	LB137374
ICSA	Lead	-19.0			-12	12	09/30/2025	14:00	LB137374
ICSAB	Lead	28.4	49.0	58	37	61	09/30/2025	14:04	LB137374
ICSA01	Lead	0.44			-12	12	10/01/2025	12:53	LB137389
ICSAB01	Lead	48.5	49.0	99	37	61	10/01/2025	12:57	LB137389
ICSA	Lead	2.04			-12	12	10/01/2025	13:02	LB137389
ICSAB	Lead	64.1	49.0	131	37	61	10/01/2025	13:06	LB137389