

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CHA Companies, Inc.

SDG No.: Q3207

Contract: CLOU03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169911BS Lead	ug/L	5000	4530		91	80 - 120	P