

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	09/23/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	09/25/25
Client Sample ID:	BP-VPB-184-GW-380-382	SDG No.:	Q3213
Lab Sample ID:	Q3213-05	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	590 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN038030.D	1	09/26/25 13:15	10/07/25 04:22	PB169904

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.42		0.11	0.34	0.34	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		78%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.39		30 - 150		97%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.36		55 - 111		90%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.36		53 - 106		90%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.57	*	58 - 132		142%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	8870		7.804			
1146-65-2	Naphthalene-d8	26100		10.605			
15067-26-2	Acenaphthene-d10	14200		14.462			
1517-22-2	Phenanthrene-d10	28000		17.21			
1719-03-5	Chrysene-d12	20300		21.384			
1520-96-3	Perylene-d12	15100		23.709			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products