

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3213</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/29/2025 18:49</u>
Lab File ID:	<u>VX047894.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.585		12.93	50
Chloromethane	0.680	0.697	0.1	2.5	50
Vinyl Chloride	0.666	0.724		8.71	50
Bromomethane	0.422	0.458		8.53	50
Chloroethane	0.445	0.479		7.64	50
Trichlorofluoromethane	0.989	1.060		7.18	50
1,1,2-Trichlorotrifluoroethane	0.578	0.619		7.09	50
1,1-Dichloroethene	0.594	0.627		5.56	50
Acetone	0.346	0.297		-14.16	50
Carbon Disulfide	1.626	1.653		1.66	50
Methyl tert-butyl Ether	2.169	2.290		5.58	50
Methyl Acetate	0.770	0.960		24.67	50
Methylene Chloride	0.732	0.780		6.56	50
trans-1,2-Dichloroethene	0.640	0.676		5.63	50
1,1-Dichloroethane	1.263	1.352	0.1	7.05	50
Cyclohexane	1.052	1.001		-4.85	50
2-Butanone	0.432	0.454		5.09	50
Carbon Tetrachloride	0.532	0.535		0.56	50
cis-1,2-Dichloroethene	0.783	0.836		6.77	50
Bromochloromethane	0.551	0.601		9.07	50
Chloroform	1.276	1.393		9.17	50
1,1,1-Trichloroethane	1.082	1.138		5.18	50
Methylcyclohexane	0.556	0.525		-5.58	50
Benzene	1.488	1.545		3.83	50
1,2-Dichloroethane	0.566	0.573		1.24	50
Trichloroethene	0.360	0.370		2.78	50
1,2-Dichloropropane	0.381	0.404		6.04	50
Bromodichloromethane	0.572	0.605		5.77	50
4-Methyl-2-Pentanone	0.477	0.529		10.9	50
Toluene	0.917	0.940		2.51	50
t-1,3-Dichloropropene	0.584	0.573		-1.88	50
cis-1,3-Dichloropropene	0.624	0.619		-0.8	50
1,1,2-Trichloroethane	0.354	0.380		7.34	50
2-Hexanone	0.343	0.362		5.54	50
Dibromochloromethane	0.407	0.447		9.83	50
1,2-Dibromoethane	0.360	0.395		9.72	50
Tetrachloroethene	0.326	0.328		0.61	50
Chlorobenzene	1.136	1.164	0.3	2.46	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3213</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/29/2025 18:49</u>
Lab File ID:	<u>VX047894.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	1.962		0.1	50
m/p-Xylenes	0.729	0.744		2.06	50
o-Xylene	0.706	0.720		1.98	50
Styrene	1.236	1.282		3.72	50
Bromoform	0.293	0.316	0.1	7.85	50
Isopropylbenzene	3.801	3.606		-5.13	50
1,1,2,2-Tetrachloroethane	1.150	1.192	0.3	3.65	50
1,3-Dichlorobenzene	1.739	1.659		-4.6	50
1,4-Dichlorobenzene	1.785	1.686		-5.55	50
1,2-Dichlorobenzene	1.657	1.607		-3.02	50
1,2-Dibromo-3-Chloropropane	0.233	0.224		-3.86	50
1,2,4-Trichlorobenzene	1.077	0.969		-10.03	50
1,2,3-Trichlorobenzene	1.002	0.936		-6.59	50
1,2-Dichloroethane-d4	0.796	0.803		0.88	50
Dibromofluoromethane	0.335	0.348		3.88	50
Toluene-d8	1.160	1.168		0.69	50
4-Bromofluorobenzene	0.440	0.447		1.59	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.