

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3213</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/30/2025 09:20</u>
Lab File ID:	<u>VX047896.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.609		17.57	20
Chloromethane	0.680	0.711	0.1	4.56	20
Vinyl Chloride	0.666	0.732		9.91	20
Bromomethane	0.422	0.433		2.61	20
Chloroethane	0.445	0.494		11.01	20
Trichlorofluoromethane	0.989	1.083		9.51	20
1,1,2-Trichlorotrifluoroethane	0.578	0.633		9.52	20
1,1-Dichloroethene	0.594	0.627		5.56	20
Acetone	0.346	0.361		4.34	20
Carbon Disulfide	1.626	1.668		2.58	20
Methyl tert-butyl Ether	2.169	2.294		5.76	20
Methyl Acetate	0.770	0.908		17.92	20
Methylene Chloride	0.732	0.791		8.06	20
trans-1,2-Dichloroethene	0.640	0.679		6.09	20
1,1-Dichloroethane	1.263	1.347	0.1	6.57	20
Cyclohexane	1.052	0.974		-7.41	20
2-Butanone	0.432	0.464		7.41	20
Carbon Tetrachloride	0.532	0.552		3.76	20
cis-1,2-Dichloroethene	0.783	0.830		6	20
Bromochloromethane	0.551	0.555		0.73	20
Chloroform	1.276	1.360		6.58	20
1,1,1-Trichloroethane	1.082	1.123		3.79	20
Methylcyclohexane	0.556	0.513		-7.73	20
Benzene	1.488	1.565		5.18	20
1,2-Dichloroethane	0.566	0.594		4.95	20
Trichloroethene	0.360	0.371		3.06	20
1,2-Dichloropropane	0.381	0.413		8.4	20
Bromodichloromethane	0.572	0.622		8.74	20
4-Methyl-2-Pentanone	0.477	0.524		9.85	20
Toluene	0.917	0.954		4.03	20
t-1,3-Dichloropropene	0.584	0.609		4.28	20
cis-1,3-Dichloropropene	0.624	0.659		5.61	20
1,1,2-Trichloroethane	0.354	0.384		8.48	20
2-Hexanone	0.343	0.366		6.71	20
Dibromochloromethane	0.407	0.452		11.06	20
1,2-Dibromoethane	0.360	0.395		9.72	20
Tetrachloroethene	0.326	0.323		-0.92	20
Chlorobenzene	1.136	1.183	0.3	4.14	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3213</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>09/30/2025</u> <u>09:20</u>
Lab File ID:	<u>VX047896.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	1.981		1.07	20
m/p-Xylenes	0.729	0.744		2.06	20
o-Xylene	0.706	0.723		2.41	20
Styrene	1.236	1.283		3.8	20
Bromoform	0.293	0.323	0.1	10.24	20
Isopropylbenzene	3.801	3.742		-1.55	20
1,1,2,2-Tetrachloroethane	1.150	1.226	0.3	6.61	20
1,3-Dichlorobenzene	1.739	1.708		-1.78	20
1,4-Dichlorobenzene	1.785	1.732		-2.97	20
1,2-Dichlorobenzene	1.657	1.671		0.85	20
1,2-Dibromo-3-Chloropropane	0.233	0.236		1.29	20
1,2,4-Trichlorobenzene	1.077	0.993		-7.8	20
1,2,3-Trichlorobenzene	1.002	0.950		-5.19	20
1,2-Dichloroethane-d4	0.796	0.783		-1.63	20
Dibromofluoromethane	0.335	0.359		7.16	20
Toluene-d8	1.160	1.184		2.07	20
4-Bromofluorobenzene	0.440	0.438		-0.46	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.