

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG NO.: Q3213
 Lab File ID: VX047868.D Date Analyzed: 09/29/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:59
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	185893	5.53	309770	6.74	268975	10.04
UPPER LIMIT	371786	6.031	619540	7.238	537950	10.537
LOWER LIMIT	92946.5	5.031	154885	6.238	134488	9.537
EPA SAMPLE NO.						
BP-VPB-184-TB-20250923	141653	5.54	299169	6.75	307943	10.04
VPB184-HYD-20250924	137990	5.54	287753	6.75	302814	10.04
BP-VPB-184-GW-340-342	121262	5.54	250493	6.75	252342	10.04
BP-VPB-184-GW-380-382	133701	5.54	285495	6.75	296119	10.04
BP-VPB-184-GW-400-402	187752	5.54	393964	6.75	400100	10.04
BP-VPB-184-GW-420-422	128773	5.54	265004	6.75	273028	10.04
BP-VPB-184-GW-440-442	126934	5.55	264916	6.75	273054	10.04
VX0929WBL01	151039	5.54	304108	6.75	305847	10.04
VX0929WBS01	165805	5.54	286774	6.75	261167	10.04
VX0929WBSD01	159091	5.54	281426	6.75	258238	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.