

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG NO.: Q3213
 Lab File ID: VX047896.D Date Analyzed: 09/30/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:20
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	159265	5.53	274684	6.74	247069	10.04
UPPER LIMIT	318530	6.031	549368	7.239	494138	10.537
LOWER LIMIT	79632.5	5.031	137342	6.239	123535	9.537
EPA SAMPLE NO.						
BP-VPB-184-GW-360-362	131080	5.54	265322	6.75	269226	10.04
BP-VPB-184-EB-20250923	127478	5.54	257240	6.75	260569	10.04
VX0930WBL01	126165	5.54	254131	6.75	265110	10.04
VX0930WBS01	139945	5.54	252162	6.74	231700	10.04

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.