



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Project: 507 Franklin Ave Nutley

RunNo.: LB137360

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	09/30/2025
Sample ID: Corrosivity	CCV1	pH	2.02	2.00	101	90-110	09/30/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	09/30/2025

Initial and Continuing Calibration Verification

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Project: 507 Franklin Ave Nutley

RunNo.: LB137366

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.096	0.099	97	85-115	09/30/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	09/30/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	09/30/2025