



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Sciaccia General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	83	70 - 130	CV	09/30/2025	09:55	LB137359
CRI01	Arsenic	22.3	20.0	112	65 - 135	P	09/30/2025	13:39	LB137374
	Barium	101	100	101	65 - 135	P	09/30/2025	13:39	LB137374
	Cadmium	5.81	6.0	97	65 - 135	P	09/30/2025	13:39	LB137374
	Chromium	9.66	10.0	97	65 - 135	P	09/30/2025	13:39	LB137374
	Lead	13.3	12.0	111	65 - 135	P	09/30/2025	13:39	LB137374
	Selenium	22.6	20.0	113	65 - 135	P	09/30/2025	13:39	LB137374
	Silver	10.1	10.0	101	65 - 135	P	09/30/2025	13:39	LB137374