



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB51	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	09:49	LB137359

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciaccia General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB57	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	09:53	LB137359
CCB58	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	10:21	LB137359
CCB59	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	10:54	LB137359
CCB60	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	11:13	LB137359
CCB61	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	11:40	LB137359

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	13:32	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	13:32	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	13:32	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	13:32	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	13:32	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	13:32	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	13:32	LB137374

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	14:18	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	14:18	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	14:18	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	14:18	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	14:18	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	14:18	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	14:18	LB137374
CCB02	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	15:50	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	15:50	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	15:50	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	15:50	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	15:50	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	15:50	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	15:50	LB137374
CCB03	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	16:53	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	16:53	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	16:53	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	16:53	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	16:53	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	16:53	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	16:53	LB137374
CCB04	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	17:43	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	17:43	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	17:43	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	17:43	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	17:43	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	17:43	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	17:43	LB137374
CCB05	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	18:59	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	18:59	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	18:59	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	18:59	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	18:59	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	18:59	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	18:59	LB137374
CCB06	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	19:49	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	19:49	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	19:49	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	19:49	LB137374

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	09/30/2025	19:49	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	19:49	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	19:49	LB137374
CCB07	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	20:39	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	20:39	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	20:39	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	20:39	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	20:39	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	20:39	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	20:39	LB137374
CCB08	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	21:29	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	21:29	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	21:29	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	21:29	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	21:29	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	21:29	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	21:29	LB137374
CCB09	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	22:18	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	22:18	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	22:18	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	22:18	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	22:18	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	22:18	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	22:18	LB137374
CCB10	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	23:04	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	23:04	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	23:04	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	23:04	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	23:04	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	23:04	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	23:04	LB137374