

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Sciacca General Contractors, LLC</u>	SDG No.: <u>Q3215</u>
Contract: <u>SCIA01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	5.01			-20	20	09/30/2025	13:44	LB137374
	Barium	7.63	6.0	127	-94	106	09/30/2025	13:44	LB137374
	Cadmium	-2.07	1.0	207	-5	7	09/30/2025	13:44	LB137374
	Chromium	49.0	52.0	94	42	62	09/30/2025	13:44	LB137374
	Lead	-6.58			-12	12	09/30/2025	13:44	LB137374
	Selenium	5.44			-20	20	09/30/2025	13:44	LB137374
	Silver	-0.092			-10	10	09/30/2025	13:44	LB137374
ICSAB01	Arsenic	111	104	107	88.4	120	09/30/2025	13:56	LB137374
	Barium	496	537	92	437	637	09/30/2025	13:56	LB137374
	Cadmium	980	972	101	826	1120	09/30/2025	13:56	LB137374
	Chromium	550	542	102	460	624	09/30/2025	13:56	LB137374
	Lead	38.9	49.0	79	37	61	09/30/2025	13:56	LB137374
	Selenium	56.7	46.0	123	26	66	09/30/2025	13:56	LB137374
	Silver	206	201	102	170	232	09/30/2025	13:56	LB137374