

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169892BS							
Arsenic	mg/Kg	40.0	34.2		86	80 - 120	P
Barium	mg/Kg	10.0	8.42		84	80 - 120	P
Cadmium	mg/Kg	10.0	8.24		82	80 - 120	P
Chromium	mg/Kg	20.0	16.9		84	80 - 120	P
Lead	mg/Kg	50.0	41.1		82	80 - 120	P
Selenium	mg/Kg	100	88.7		89	80 - 120	P
Silver	mg/Kg	3.8	3.28		86	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169897BS Mercury	mg/Kg	0.25	0.27		106	80 - 120	CV