



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB52	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	14:18	LB137368

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciaccia General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB62	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	14:22	LB137368
CCB63	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	14:50	LB137368
CCB64	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	15:17	LB137368
CCB65	Mercury	0.076	+/-0.2	U	0.20	CV	09/30/2025	15:45	LB137368

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	13:32	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	13:32	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	13:32	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	13:32	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	13:32	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	13:32	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	13:32	LB137374

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	14:18	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	14:18	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	14:18	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	14:18	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	14:18	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	14:18	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	14:18	LB137374
CCB02	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	15:50	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	15:50	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	15:50	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	15:50	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	15:50	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	15:50	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	15:50	LB137374
CCB03	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	16:53	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	16:53	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	16:53	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	16:53	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	16:53	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	16:53	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	16:53	LB137374
CCB04	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	17:43	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	17:43	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	17:43	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	17:43	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	17:43	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	17:43	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	17:43	LB137374
CCB05	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	18:59	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	18:59	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	18:59	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	18:59	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	18:59	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	18:59	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	18:59	LB137374
CCB06	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	19:49	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	19:49	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	19:49	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	19:49	LB137374

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	09/30/2025	19:49	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	19:49	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	19:49	LB137374
CCB07	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	20:39	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	20:39	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	20:39	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	20:39	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	20:39	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	20:39	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	20:39	LB137374
CCB08	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	21:29	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	21:29	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	21:29	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	21:29	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	21:29	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	21:29	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	21:29	LB137374
CCB09	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	22:18	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	22:18	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	22:18	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	22:18	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	22:18	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	22:18	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	22:18	LB137374
CCB10	Arsenic	5.12	+/-10	U	20.0	P	09/30/2025	23:04	LB137374
	Barium	14.6	+/-50	U	100	P	09/30/2025	23:04	LB137374
	Cadmium	0.50	+/-3	U	6.00	P	09/30/2025	23:04	LB137374
	Chromium	2.12	+/-5	U	10.0	P	09/30/2025	23:04	LB137374
	Lead	2.30	+/-6	U	12.0	P	09/30/2025	23:04	LB137374
	Selenium	9.64	+/-10	U	20.0	P	09/30/2025	23:04	LB137374
	Silver	1.62	+/-5	U	10.0	P	09/30/2025	23:04	LB137374

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciaccia General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	12:45	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	12:45	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	12:45	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	12:45	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	12:45	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	12:45	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	12:45	LB137389

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	13:14	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	13:14	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	13:14	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	13:14	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	13:14	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	13:14	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	13:14	LB137389
CCB02	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	14:06	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	14:06	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	14:06	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	14:06	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	14:06	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	14:06	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	14:06	LB137389
CCB03	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	15:04	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	15:04	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	15:04	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	15:04	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	15:04	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	15:04	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	15:04	LB137389
CCB04	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	15:55	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	15:55	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	15:55	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	15:55	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	15:55	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	15:55	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	15:55	LB137389
CCB05	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	16:46	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	16:46	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	16:46	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	16:46	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	16:46	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	16:46	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	16:46	LB137389
CCB06	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	17:41	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	17:41	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	17:41	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	17:41	LB137389

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	10/01/2025	17:41	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	17:41	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	17:41	LB137389
CCB07	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	18:32	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	18:32	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	18:32	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	18:32	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	18:32	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	18:32	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	18:32	LB137389
CCB08	Arsenic	5.12	+/-10	U	20.0	P	10/01/2025	19:23	LB137389
	Barium	14.6	+/-50	U	100	P	10/01/2025	19:23	LB137389
	Cadmium	0.50	+/-3	U	6.00	P	10/01/2025	19:23	LB137389
	Chromium	2.12	+/-5	U	10.0	P	10/01/2025	19:23	LB137389
	Lead	2.30	+/-6	U	12.0	P	10/01/2025	19:23	LB137389
	Selenium	9.64	+/-10	U	20.0	P	10/01/2025	19:23	LB137389
	Silver	1.62	+/-5	U	10.0	P	10/01/2025	19:23	LB137389

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Sciacca General Contractors, LLC

SDG No.: Q3215

Contract: SCIA01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------