

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q3215

Continuing Calib Date: 09/30/2025

Initial Calibration Date(s): 09/15/2025

09/15/2025

Continuing Calib Time: 16:41

Initial Calibration Time(s): 09:34

19:21

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.54       | 5.55      | 5.45      | 5.65 | 0.01       |
| Aroclor-1254-2 (2)   | 5.69       | 5.70      | 5.60      | 5.80 | 0.01       |
| Aroclor-1254-3 (3)   | 6.09       | 6.10      | 6.00      | 6.20 | 0.01       |
| Aroclor-1254-4 (4)   | 6.32       | 6.33      | 6.23      | 6.43 | 0.01       |
| Aroclor-1254-5 (5)   | 6.74       | 6.75      | 6.65      | 6.85 | 0.01       |
| Tetrachloro-m-xylene | 3.66       | 3.66      | 3.56      | 3.76 | 0.00       |
| Decachlorobiphenyl   | 8.66       | 8.67      | 8.57      | 8.77 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q3215

Continuing Calib Date: 09/30/2025

Initial Calibration Date(s): 09/15/2025

09/15/2025

Continuing Calib Time: 16:41

Initial Calibration Time(s): 09:34

19:21

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.51       | 5.52      | 5.42      | 5.62 | 0.01       |
| Aroclor-1254-2 (2)   | 5.66       | 5.67      | 5.57      | 5.77 | 0.01       |
| Aroclor-1254-3 (3)   | 6.06       | 6.07      | 5.97      | 6.17 | 0.01       |
| Aroclor-1254-4 (4)   | 6.29       | 6.29      | 6.19      | 6.39 | 0.00       |
| Aroclor-1254-5 (5)   | 6.70       | 6.71      | 6.61      | 6.81 | 0.01       |
| Tetrachloro-m-xylene | 3.65       | 3.65      | 3.55      | 3.75 | 0.00       |
| Decachlorobiphenyl   | 8.61       | 8.62      | 8.52      | 8.72 | 0.01       |