

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q3215

Continuing Calib Date: 09/30/2025

Initial Calibration Date(s): 09/15/2025

09/15/2025

Continuing Calib Time: 21:55

Initial Calibration Time(s): 09:34

19:21

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.74	4.74	4.64	4.84	0.00
Aroclor-1248-2 (2)	4.97	4.98	4.88	5.08	0.01
Aroclor-1248-3 (3)	5.19	5.19	5.09	5.29	0.00
Aroclor-1248-4 (4)	5.54	5.55	5.45	5.65	0.01
Aroclor-1248-5 (5)	5.58	5.59	5.49	5.69	0.01
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.01
Decachlorobiphenyl	8.66	8.67	8.57	8.77	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q3215

Continuing Calib Date: 09/30/2025

Initial Calibration Date(s): 09/15/2025

09/15/2025

Continuing Calib Time: 21:55

Initial Calibration Time(s): 09:34

19:21

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.72	4.72	4.62	4.82	0.00
Aroclor-1248-2 (2)	4.95	4.96	4.86	5.06	0.01
Aroclor-1248-3 (3)	4.99	5.00	4.90	5.10	0.01
Aroclor-1248-4 (4)	5.16	5.17	5.07	5.27	0.01
Aroclor-1248-5 (5)	5.55	5.56	5.46	5.66	0.01
Tetrachloro-m-xylene	3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl	8.61	8.61	8.51	8.71	0.00