

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q3215

Continuing Calib Date: 10/03/2025

Initial Calibration Date(s): 09/15/2025

09/15/2025

Continuing Calib Time: 15:54

Initial Calibration Time(s): 09:34

19:21

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.74	4.74	4.64	4.84	0.00
Aroclor-1242-2 (2)	4.76	4.76	4.66	4.86	0.00
Aroclor-1242-3 (3)	4.81	4.82	4.72	4.92	0.01
Aroclor-1242-4 (4)	4.93	4.94	4.84	5.04	0.01
Aroclor-1242-5 (5)	5.58	5.59	5.49	5.69	0.01
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl	8.66	8.67	8.57	8.77	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: SCIA01

Lab Code: ACE

SDG NO.: Q3215

Continuing Calib Date: 10/03/2025

Initial Calibration Date(s): 09/15/2025

09/15/2025

Continuing Calib Time: 15:54

Initial Calibration Time(s): 09:34

19:21

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.72	4.72	4.62	4.82	0.00
Aroclor-1242-2 (2)	4.73	4.74	4.64	4.84	0.01
Aroclor-1242-3 (3)	4.91	4.92	4.82	5.02	0.01
Aroclor-1242-4 (4)	4.99	5.00	4.90	5.10	0.01
Aroclor-1242-5 (5)	5.51	5.52	5.42	5.62	0.01
Tetrachloro-m-xylene	3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl	8.60	8.62	8.52	8.72	0.02