

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCIA01
Lab Code: ACE **SDG NO.:** Q3215
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114092.D **Time Analyzed:** 13:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.740	4.643	4.843	483.610	500.000	-3.3
Aroclor-1016-2	4.758	4.661	4.861	497.150	500.000	-0.6
Aroclor-1016-3	4.815	4.718	4.918	478.990	500.000	-4.2
Aroclor-1016-4	4.934	4.837	5.037	477.400	500.000	-4.5
Aroclor-1016-5	5.190	5.093	5.293	526.280	500.000	5.3
Aroclor-1260-1	6.224	6.129	6.329	520.870	500.000	4.2
Aroclor-1260-2	6.413	6.319	6.519	504.340	500.000	0.9
Aroclor-1260-3	6.778	6.683	6.883	511.060	500.000	2.2
Aroclor-1260-4	7.036	6.941	7.141	564.800	500.000	13.0
Aroclor-1260-5	7.280	7.185	7.385	584.450	500.000	16.9
Decachlorobiphenyl	8.663	8.570	8.770	53.090	50.000	6.2
Tetrachloro-m-xylene	3.658	3.560	3.760	50.960	50.000	1.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCIA01
Lab Code: ACE **SDG NO.:** Q3215
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114092.D **Time Analyzed:** 13:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.623	4.823	455.410	500.000	-8.9
Aroclor-1016-2	4.737	4.641	4.841	459.530	500.000	-8.1
Aroclor-1016-3	4.912	4.816	5.016	460.700	500.000	-7.9
Aroclor-1016-4	4.955	4.858	5.058	463.200	500.000	-7.4
Aroclor-1016-5	5.166	5.070	5.270	477.210	500.000	-4.6
Aroclor-1260-1	6.191	6.096	6.296	445.220	500.000	-11.0
Aroclor-1260-2	6.379	6.284	6.484	428.690	500.000	-14.3
Aroclor-1260-3	6.529	6.435	6.635	416.750	500.000	-16.7
Aroclor-1260-4	6.998	6.903	7.103	428.270	500.000	-14.3
Aroclor-1260-5	7.240	7.146	7.346	423.670	500.000	-15.3
Decachlorobiphenyl	8.607	8.514	8.714	49.120	50.000	-1.8
Tetrachloro-m-xylene	3.649	3.551	3.751	48.090	50.000	-3.8