

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCIA01
Lab Code: ACE **SDG NO.:** Q3215
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114107.D **Time Analyzed:** 00:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.737	4.643	4.843	495.210	500.000	-1.0
Aroclor-1016-2	4.756	4.661	4.861	487.970	500.000	-2.4
Aroclor-1016-3	4.812	4.718	4.918	478.930	500.000	-4.2
Aroclor-1016-4	4.931	4.837	5.037	474.780	500.000	-5.0
Aroclor-1016-5	5.188	5.093	5.293	555.880	500.000	11.2
Aroclor-1260-1	6.222	6.129	6.329	495.220	500.000	-1.0
Aroclor-1260-2	6.412	6.319	6.519	476.350	500.000	-4.7
Aroclor-1260-3	6.776	6.683	6.883	420.340	500.000	-15.9
Aroclor-1260-4	7.035	6.941	7.141	434.630	500.000	-13.1
Aroclor-1260-5	7.279	7.185	7.385	436.460	500.000	-12.7
Decachlorobiphenyl	8.662	8.570	8.770	39.250	50.000	-21.5
Tetrachloro-m-xylene	3.657	3.560	3.760	50.160	50.000	0.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCIA01
Lab Code: ACE **SDG NO.:** Q3215
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114107.D **Time Analyzed:** 00:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.718	4.623	4.823	457.450	500.000	-8.5
Aroclor-1016-2	4.737	4.641	4.841	455.250	500.000	-9.0
Aroclor-1016-3	4.911	4.816	5.016	469.510	500.000	-6.1
Aroclor-1016-4	4.954	4.858	5.058	463.760	500.000	-7.2
Aroclor-1016-5	5.165	5.070	5.270	496.290	500.000	-0.7
Aroclor-1260-1	6.190	6.096	6.296	422.390	500.000	-15.5
Aroclor-1260-2	6.379	6.284	6.484	427.980	500.000	-14.4
Aroclor-1260-3	6.529	6.435	6.635	386.400	500.000	-22.7
Aroclor-1260-4	6.997	6.903	7.103	388.080	500.000	-22.4
Aroclor-1260-5	7.241	7.146	7.346	382.020	500.000	-23.6
Decachlorobiphenyl	8.607	8.514	8.714	42.750	50.000	-14.5
Tetrachloro-m-xylene	3.649	3.551	3.751	47.680	50.000	-4.6