

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCIA01
Lab Code: ACE **SDG NO.:** Q3215
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114131.D **Time Analyzed:** 11:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.739	4.643	4.843	492.100	500.000	-1.6
Aroclor-1016-2	4.757	4.661	4.861	499.930	500.000	0.0
Aroclor-1016-3	4.814	4.718	4.918	485.240	500.000	-3.0
Aroclor-1016-4	4.932	4.837	5.037	481.290	500.000	-3.7
Aroclor-1016-5	5.189	5.093	5.293	516.920	500.000	3.4
Aroclor-1260-1	6.223	6.129	6.329	538.200	500.000	7.6
Aroclor-1260-2	6.413	6.319	6.519	486.450	500.000	-2.7
Aroclor-1260-3	6.778	6.683	6.883	464.880	500.000	-7.0
Aroclor-1260-4	7.036	6.941	7.141	567.910	500.000	13.6
Aroclor-1260-5	7.279	7.185	7.385	574.910	500.000	15.0
Decachlorobiphenyl	8.663	8.570	8.770	47.190	50.000	-5.6
Tetrachloro-m-xylene	3.658	3.560	3.760	51.950	50.000	3.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** SCIA01
Lab Code: ACE **SDG NO.:** Q3215
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114131.D **Time Analyzed:** 11:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.718	4.623	4.823	446.720	500.000	-10.7
Aroclor-1016-2	4.735	4.641	4.841	456.120	500.000	-8.8
Aroclor-1016-3	4.910	4.816	5.016	451.670	500.000	-9.7
Aroclor-1016-4	4.953	4.858	5.058	457.810	500.000	-8.4
Aroclor-1016-5	5.164	5.070	5.270	470.120	500.000	-6.0
Aroclor-1260-1	6.190	6.096	6.296	441.170	500.000	-11.8
Aroclor-1260-2	6.378	6.284	6.484	420.410	500.000	-15.9
Aroclor-1260-3	6.528	6.435	6.635	407.890	500.000	-18.4
Aroclor-1260-4	6.997	6.903	7.103	406.950	500.000	-18.6
Aroclor-1260-5	7.240	7.146	7.346	404.380	500.000	-19.1
Decachlorobiphenyl	8.606	8.514	8.714	48.290	50.000	-3.4
Tetrachloro-m-xylene	3.648	3.551	3.751	48.320	50.000	-3.4